



Asbestos in Materials Individual Results : Round 079 : 079 AIMS R79

For Laboratory Number : 1367 Scientific Services Ltd (Cornwall)

Report Issued 20/06/2023 13:43:15


Report No. 079/187/16599/ 25226

Your Performance Score 0

Your Performance Classification Good

Your Score This Round 0

Sample	Your Result	Assigned Result	Sample Score
1	No Asbestos	No Asbestos	0
2	Amosite, Tremolite	Amosite, Anthophyllite	0
3	No Asbestos	No Asbestos	0
4	Crocidolite	Crocidolite	0



This report is confidential to your laboratory and can only be accessed by using your unique username and password. Once printed, the proficiency testing provider can not be held responsible for its confidentiality.

HSE
Melanie Clunas
 AIMS Scheme Co-Ordinator
 Fibres Team
 Final Report

HSE, Science Division
 Harpur Hill, Buxton, Derbyshire, SK17 9JN - UK

FTPT F0903 **ISSUE 5** Page 1 of 1

