

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 2577 Accredited to ISO/IEC 17025:2017	Scientific Services Ltd	
	Issue No: 022 Issue date: 01 February 2022	
	The Fuel Depot Scorrier Redruth Cornwall TR16 5UT	Contact: Dr T Hamilton Tel: +44 (0)1209 311350 E-Mail: t.hamilton@scientificservices.org Website: www.scientificservices.org
Testing performed by the Organisation at the locations specified below		

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code	
Address The Fuel Depot Scorrier Redruth Cornwall TR16 5UT	Local contact Dr T Hamilton Tel: +44 (0)1209 311350 Email: t.hamilton@scientificservices.org	Health and Hygiene Head Office Asbestos – All Support Functions	A

Site activities performed away from the locations listed above:

Location details	Activity	Location code
Client Premises	Health and Hygiene	B
Mobile Testing Laboratories	Health and Hygiene	C



2577
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Scientific Services Ltd

Issue No: 022 **Issue date:** 01 February 2022

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ASBESTOS FIBRES IN AIR	<u>Health and Hygiene</u>	Health and Safety Executive - Asbestos: The Analysts' Guide (HSG 248) – 2021	
	Sampling of air for fibre counting	Documented In-House Method D3TA2 based on HSG 248	B, C
	Fibre counting	Documented In-House Method D3TA2, Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	A, B, C
ASBESTOS IN BULK MATERIALS including materials and products suspected of containing asbestos	4 Stage Clearance Process	Documented In-House Method D3TA3, Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	B
	Sampling of bulk materials for subsequent identification of asbestos	Documented In-House Method D3TA8 based on HSG 248	B
	Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	Documented In-House Method D3TA1 using stereo-microscopy, polarised light optical microscopy and dispersion staining based on HSG 248	A

END