



Asbestos in Materials Individual Results : Round 087 : 087 AIMS R87

For Laboratory Number : 1367 Scientific Services Ltd (Cornwall)

Report Issued 04/03/2026 11:00:47


Report No. 087/187/20487/ 40051

Your Performance Score 0

Your Performance Classification Good

Your Score This Round 0

Sample	Your Result	Assigned Result	Sample Score
1	No Asbestos	No Asbestos	0
2	Chrysotile	Chrysotile	0
3	Amosite	Amosite	0
4	No Asbestos	No Asbestos	0



This report is confidential to your laboratory and can only be accessed by using your unique username and password. Once printed, the proficiency testing provider can not be held responsible for its confidentiality.

Melanie Clunas  
AIMS Scheme Co-Ordinator  
FIBRES & MINERALS TEAM

HSE, Science Division  
Harpur Hill, Buxton, Derbyshire, SK17 9JN - UK

